

6390 Joyce Drive
100
Golden, CO 80403

Phone 303-940-0033
Fax 866-283-0269
www.wibby.com

September 7, 2010

Larry Lem
Calscience Environmental Laboratories, Inc.
7440 Lincoln Way
Garden Grove, CA 92841-1427
United States

Dear Larry,

Thank you for participating in the RR-06974 Rapid Return™ Proficiency Testing Study. Enclosed is your final report, which has been carefully reviewed by the PT specialists at Wibby Environmental. A final report for your laboratory has been sent to all accrediting agencies you requested at the time of your data submittal.

Thank you again for participating in the RR-06974 Rapid Return™ Proficiency Testing Study. We appreciate working with you and look forward to our next study. If you have any questions, please call Chuck Wibby, President of Wibby Environmental, or me at 303-940-0033.

Sincerely,



Anne Lang
QA Manager

Report Definitions:

NELAC Experimental Analytes

NELAC Experimental Analytes must be run twice a year. Due to the experimental nature of the acceptance limits, accreditation status is not based on the evaluation received.

Assigned Value

The Assigned Value is determined from the study mean, gravimetric and volumetric true concentration of an analyte to be analyzed, calculation and/or an appropriate reference value as stipulated in the EPA National Standards for Water Proficiency Testing Studies Criteria Document (current version), the National Environmental Laboratory Accreditation Conference (NELAC) criteria (ref: NELAC FOT tables, NELAC PT Committee) and other documents distributed by accrediting agencies as applicable.

Evaluation Limits

Acceptance Limits are derived from fixed limits, coefficients, constants and calculations stipulated in the EPA National Standards for Water Proficiency Testing Studies Criteria Documents (current version), the National Environmental Laboratory Accreditation Conference (NELAC) criteria (ref: NELAC PT FOT tables, NELAC PT Committee) and other documents distributed by accrediting agencies as applicable.

Evaluation

<i>Acceptable</i>	The reported value falls within the Acceptance Limits.
<i>Not Acceptable</i>	The reported value falls outside the Acceptance Limits.
<i>No Evaluation</i>	The reported value is non-numeric and can not be evaluated.
<i>NR</i>	As required by the NELAC standards and requested by state authorities, any analyte purchased but not reported by your facility is listed as NR (Not Reported).

6390 Joyce Drive
100
Golden, CO 80403

Phone 303-940-0033
Fax 866-283-0269
www.wibby.com

State Regulatory Agency Request(s):

Wibby Environmental has been authorized to send a copy of your RR-06974 final results to the following state agencies:

CA, CAELAP, NV, OR and WA

Other Regulatory Agency Request(s):

No requests were made to send a copy of your RR-06974 final results to an agency.

Third Party Request(s):

Wibby Environmental has been requested to send a copy of your RR-06974 final results to the following third parties:

Perry Johnson Laboratory Accreditation, Inc.
755 W. Big Beaver Road, Suite 1325
Troy, MI 48084
Tracy Szerszen
pjlabs@pjlabs.com

6390 Joyce Drive
 # 100
 Golden, CO 80403

Phone 303-940-0033
 Fax 866-283-0269
 www.wibby.com

Final Report - Rapid Return™

Study: RR-06974

Opening Date: August 5, 2010 - Closing Date: September 1, 2010

Laboratory: Calscience Environmental Laboratories, Inc.
 7440 Lincoln Way
 Garden Grove, CA 92841-1427
 UNITED STATES

Contact: Larry Lem
 714-895-5494

EPA Lab ID: CA00111

Ignitability of Solids (RR-IGN-SOIL)							Lot #: RR-06974-82		
NELAC Code	Analyte	Method Code	Method Description	Units	Assigned Value	Result	Acceptance Limits	Evaluation	
Additional State Specific Analytes									
-	Burn Rate		EPA 1030 (Jar #1)	mm/sec	1.39	1.14	0.793 - 1.98	Acceptable	
-	Burn Rate		EPA 1030 (Jar #2)	mm/sec	1.39	1.19	0.793 - 1.98	Acceptable	